

22	High definition TV	1
(C)	Postoperative Recovery Room	
23	Oxygen therapy unit	8
24	Blood Pressure Monitor (digital/Electronic)	8
25	Pulse oximeter	8
26	E.C.G. Monitor	8
27	Suction Machine	8
28	Provision for resuscitation equipment and CPR Algorithms	1
29	Airway crash cart	1
30	ABG machine (Optional)	1
31	Side lab for emergency investigations (Optional)	1
32	TOF monitor (PNS) (Optional)	1
(D)	Pain Clinic	
33	Radio frequency ablation machine	2
34	Fluoroscopy machine (e-ARM)(may be shared with other department)	2
35	Ultrasound machine	1
36	Patients controlled analgesia system (portable)	1
37	Styleted Epidural catheter	10
38	Rac's catheters	As required
39	Intrathecal infusion pumps	As required
40	Pain assessment chart	4
41	OT Table fluoroscopy compatible	2
42	Nerve locator	1
43	Computer	2
44	Anodyne Machine (Optional) (For diabetic foot care)	2
45	Transcutaneous Electric Nerve Stimulating Machine (optional)	1
46	Anesthetic machine with resuscitation equipment	2
47	Crash cart	2
48	Monitors for vital signs (NIBP, P.R, ECG, SPO2, Temperature, R.R)	2
49	Nerve stimulator	2
50	18 G epidural needle	4
51	Multimedia Projector with screen	1

20. RADIO-DIAGNOSIS (New List)

S.No.	Name of the Items	No. Required
(A)	General	
1	Conventional X-ray Unit for routine X ray 300 mA, 500mA, 800mA (with IITV) - 1 each	2+2+1 Total - 5
2	CR system	2
3.	DR System	-
(B)	Mobile X-ray units	
3	a) 60 mA	3
4	b) 100mA	3

5	Ultrasonography equipment and color Doppler (Additional unit independently for OBGY).	3+1
6	CT (16 slice).	1
7	Mammography (Preferably).	1
8	MRI (Preferably).	1
9	Multimedia Projector with Screen	1
	Note: The department should have resuscitation equipment like laryngoscope, all size endotracheal tubes, nasal and oral airways, Magill's forceps, LMA / PLMA of all sizes, Electrical suction apparatus	

21. CENTRAL CASUALTY DEPT. (New List)

S.No.	NAME OF THE ITEMS	No. Required
	There shall be a Central Casualty department wherein consultancy services shall be provided by the faculty of relevant Departments. And relevant investigation facilities to be run in liaison with central laboratory.	
1	Casualty resuscitation room should have central piped Oxygen supply & suction on every bed.	As required
2	Ventilators, Multipara monitors, ECG machines, Nebulizers.	3 each
3	Fully equipped disaster trolleys (emergency trolleys).	3
4	Emergency X-ray (Fixed 300/500mA & Mobile 100mA)	1 each
5	Sonography machine/s in the casualty	1
6	Plaster room	1
	Operation theatre shall be provided. These shall be equipped for emergency operative treatment and include the following:	
7	Operation Theatre Table	1
8	Operation Theatre Ceiling light	1
9	Pedestal lights	2
10	Electro-surgical cautery unit	2
11	Suction	2
12	Pulse oximeter	2
13	Anesthesia Equipment	1 set
14	Resuscitation kit	1
15	Assorted surgical instrument sets	5
16	Autoclave	1
17	OT should have resuscitation equipment like Anesthesia machine and accessories like laryngoscope, all size endotracheal tubes, nasal and oral airways, Magill's forceps, Mechanical ventilator Separate or with anesthesia machine, LMA / PLMA of all sizes, Electrical suction apparatus	
	Required equipment on Each Emergency Trolley	
(A)	Airway:	
18	Pocket mask with oxygen port	As required
19	Digital Blood Pressure apparatus	As required
20	Ambu bag and tubing (ideally, the resuscitation bag should be single use – if not, it should be equipped with a suitable filter)	As required
21	Clear face masks, sizes 3, 4 and 5	As required
22	Oropharyngeal airways, sizes 2, 3 and 4	As required
23	Nasopharyngeal airways, sizes 6 and 7	As required
24	Portable suction equipment	As required
25	Yankauer suckers	As required